



DIVELY SCALE CO., INC.

4012 Alameda Ave., El Paso, TX 79905-3799
 (915) 533-6712 or (888) 383-6744 FAX 533-6755
 1607 4th St. N.W., Albuquerque, NM, 87102-1622
 (505) 243-2189 FAX 242-2485

Calibration Traceability Certificate

Work Order No.: 9845Page 1 of 4

Customer:	CONVENTION CERT		Next Calibration Due	
Address:	701 2nd ST NW 8		Work Instruction: TF	N/A
City, State & Zip:	Albuquerque, NM 87102		Environmental Conditions	
Service Location (If Different):			Temperature (°F):	N/A
Calibration Date:	9/25/14		Relative Humidity (%):	N/A
Calibration Interval:	As Needed		Any Adverse Conditions (Yes/No)?	No
Indicator Mfg:	Aand A Scales LLC	Model #: 2	Serial #: A20130331061	Indicator ID: (Cust Assigned) —
Base Mfg:	—	Model #: —	Serial #: —	Base ID: (Cust Assigned) —
Capacity:	400lb	Readability: 0.02	Divisions: 20000	Class: UM
				REPEATABILITY (Pass/Fail)
				As Found: Pass
				As Left: —
TOLERANCE CODE	Handbook 44:	T.6	Customer Specified:	—
				Other (See Comments): —

Section, Corner, or Side-To-Side Test

INDICATOR	Position	As Found		Position	As Left	
		Value	Error		Value	Error
Platform Layout 2 3 4 1	1	0.00		1		
	2	0.00	Error 0.00	2		Error
	3	0.00	Tolerance +/- 0.02	3	BP	Tolerance +/-
	4	0.00	Pass / Fail Pass	4		Pass / Fail
	5	—	Load Used: 100.00	5		Load Used:
	6	—		6		

Load Test

Test Weight Applied	As Found Reading	Error	Tolerance +/-	Pass / Fail	As Left Reading	Pass / Fail
ZERO	0.00	0.00	0.00	Pass		
100.00	100.00	0.00	0.02	Pass	BP	
200.00	200.00	0.00	0.02	Pass		
300.00	300.00	0.00	0.04	Pass		
400.00	400.00	0.00	0.04	Pass		
200.00	200.00	0.00	0.02	Pass		
0.00	0.00	0.00	0.02	Pass		

SN of N.I.S.T. Traceable Test Weights Used: A60 - A65

N.I.S.T. Reference Number: 13-187

Comments:

Technician: [Signature] State Reg. No.: 195 Date: 9-25-14

FOR QUESTIONS AND CLARIFICATIONS PLEASE CONTACT THE SERVICE MANAGER

This certificate is NOT ISO 17025 compliant and should not be used as a substitute for an ISO 17025 certified report. Dively Scale Co., Inc. certifies that the above listed equipment pass or fail based on a comparison between displayed results and the applicable tolerance. Our estimate of Measurement Uncertainty is not considered in these pass / fail conclusions and can be provided upon request for an additional fee. Our test methods comply with National and International Standards (e.g. N.I.S.T. Handbook 44 where applicable) or by validated customer specified methods. The test weights used are certified to comply with standard masses, traceable to N.I.S.T. via a ISO 17025 accredited or N.I.S.T. approved labs. Any discrepancies beyond acceptable tolerances, in the aforementioned equipment, are reported to the appropriate facility's manager for authorization of further repairs and/or replacement as necessary. This report shall not be reproduced except in full without expressed written consent from Dively Scale Co., Inc.



DIVELY SCALE CO., INC.

4012 Alameda Ave., El Paso, TX 79905-3799
 (915) 533-6712 or (888) 383-6744 FAX 533-6755
 1607 4th St. N.W., Albuquerque, NM, 87102-1622
 (505) 243-2189 FAX 242-2485

Calibration Traceability Certificate

Work Order No.: 9845

Page 2 of 4

Customer: <u>CONVENTION CENTER</u>	Next Calibration Due: _____	
Address: <u>401 2ND ST NW</u>	Work Instruction: TF <u>N/A</u>	
City, State & Zip: <u>Albuquerque, NM 87102</u>	Environmental Conditions	
Service Location (If Different): _____		Temperature (°F): <u>N/A</u>
Calibration Date: <u>9-25-14</u>		Relative Humidity (%): <u>N/A</u>
Calibration Interval: <u>As needed</u>		Any Adverse Conditions (Yes/No)? <u>NO</u>
Indicator Mfg: <u>AandAScales LLC</u> Model #: <u>1</u> Serial #: <u>AC2030331109</u> Indicator ID: _____ (Cust Assigned)	REPEATABILITY	
Base Mfg: _____ Model #: _____ Serial #: _____ Base ID: _____ (Cust Assigned)	(Pass/Fail)	
Capacity: <u>4001b</u> Readability: <u>0.021b</u> Divisions: <u>20000</u> Class: <u>UM</u>	As Found: <u>Pass</u>	As Left: _____
TOLERANCE CODE Handbook 44: <u>71.1</u> Customer Specified: _____ Other (See Comments): _____		

Section, Corner, or Side-To-Side Test

INDICATOR	Position	As Found		Position	As Left	
		Value	Error		Value	Error
Platform Layout	1	0.00		1		
	2	0.00	Error 0.00	2		Error
	3	0.00	Tolerance +/- 0.01	3	<u>BP</u>	Tolerance +/-
	4	0.00	Pass / Fail <u>Pass</u>	4		Pass / Fail
	5	—	Load Used: <u>100.00</u>	5		Load Used:
	6	—		6		

Load Test

Test Weight Applied	As Found Reading	Error	Tolerance +/-	Pass / Fail	As Left Reading	Pass / Fail
ZERO	0.00	0.00	0.00	<u>Pass</u>		
100.00	100.00	0.00	0.02	<u>Pass</u>		
200.00	200.00	0.00	0.02	<u>Pass</u>	<u>BP</u>	
300.00	300.00	0.00	0.04	<u>Pass</u>		
400.00	400.00	0.00	0.04	<u>Pass</u>		
200.00	200.00	0.00	0.02	<u>Pass</u>		
0.00	0.00	0.00	0.02	<u>Pass</u>		

S/N of N.I.S.T. Traceable Test Weights Used:


N.I.S.T. Reference Number	<u>13-187</u>
---------------------------	---------------

Comments:

Technician: [Signature] State Reg. No.: 195 Date: 9-25-14

FOR QUESTIONS AND CLARIFICATIONS PLEASE CONTACT THE SERVICE MANAGER

This certificate is NOT ISO 17025 compliant and should not be used as a substitute for an ISO 17025 certified report. Dively Scale Co., Inc. certifies that the above listed equipment pass or fail based on a comparison between displayed results and the applicable tolerance. Our estimate of Measurement Uncertainty is not considered in these pass / fail conclusions and can be provided upon request for an additional fee. Our test methods comply with National and International Standards (e.g. N.I.S.T. Handbook 44 where applicable) or by validated customer specified methods. The test weights used are certified to comply with standard masses, traceable to N.I.S.T. via a ISO 17025 accredited or N.I.S.T. approved labs. Any discrepancies beyond acceptable tolerances, in the aforementioned equipment, are reported to the appropriate facility's manager for authorization of further repairs and/or replacement as necessary. This report shall not be reproduced except in full without expressed written consent from Dively Scale Co., Inc.



DIVELY SCALE CO., INC.

4012 Alameda Ave., El Paso, TX 79905-3799
 (915) 533-6712 or (888) 383-6744 FAX 533-6755
 1607 4th St. N.W., Albuquerque, NM, 87102-1622
 (505) 243-2189 FAX 242-2485

Calibration Traceability Certificate

Work Order No.: 9845Page 3 of 4

Customer:	<u>Double Tree Hotel</u>	Next Calibration Due	<u> </u>
Address:	<u>201 Marquette Ave NW</u>	Work Instruction:	TF N/A
City, State & Zip:	<u>Albuquerque, NM 87102</u>	Environmental Conditions	
Service Location (If Different):		Temperature (°F):	N/A
Calibration Date:	<u>9-25-14</u>	Relative Humidity (%):	N/A
Calibration Interval:	<u>As needed</u>	Any Adverse Conditions (Yes/No)?	<u>NO</u>
Indicator Mfg:	<u>Total Comp</u>	Model #:	<u>T500E</u>
Base Mfg:		Serial #:	<u>70201010000000000000</u>
Capacity:	<u>400 lb</u>	Readability:	<u>0.02lb</u>
		Divisions:	<u>20000</u>
		Class:	<u>UM</u>
TOLERANCE CODE		Handbook 44:	<u>T.61</u>
		Customer Specified:	<u> </u>
		Other (See Comments):	<u> </u>

REPEATABILITY	
(Pass/Fail)	
As Found:	<u>Pass</u>
As Left:	<u> </u>

Section, Corner, or Side-To-Side Test

INDICATOR	Position	As Found		Position	As Left	
		Value	Error		Value	Error
Platform Layout 3 2 4 1	1	<u>0.00</u>		1		
	2	<u>0.00</u>	Error <u>0.00</u>	2		Error
	3	<u>0.00</u>	Tolerance +/- <u>0.02</u>	3	<u>BP</u>	Tolerance +/-
	4	<u>0.00</u>	Pass / Fail <u>Pass</u>	4		Pass / Fail
	5	<u> </u>	Load Used: <u>100.00</u>	5		Load Used:
	6	<u> </u>		6		

Load Test

Test Weight Applied	As Found Reading	Error	Tolerance +/-	Pass / Fail	As Left Reading	Pass / Fail
ZERO	<u>0.00</u>	<u>0.00</u>	<u>0.00</u>	<u>Pass</u>		
<u>100.00</u>	<u>100.00</u>	<u>0.00</u>	<u>0.02</u>	<u>Pass</u>		
<u>200.00</u>	<u>200.00</u>	<u>0.00</u>	<u>0.02</u>	<u>Pass</u>	<u>BP</u>	
<u>300.00</u>	<u>300.00</u>	<u>0.00</u>	<u>0.04</u>	<u>Pass</u>		
<u>400.00</u>	<u>400.00</u>	<u>0.00</u>	<u>0.04</u>	<u>Pass</u>		
<u>200.00</u>	<u>200.00</u>	<u>0.00</u>	<u>0.02</u>	<u>Pass</u>		
<u>0.00</u>	<u>0.00</u>	<u>0.00</u>	<u>0.02</u>	<u>Pass</u>		

S/N of N.I.S.T. Traceable Test Weights Used: A60-105

N.I.S.T. Reference Number: 13-187

Comments:

Technician: BJJ. DM State Reg. No.: 195 Date: 9-25-14

FOR QUESTIONS AND CLARIFICATIONS PLEASE CONTACT THE SERVICE MANAGER

This certificate is NOT ISO 17025 compliant and should not be used as a substitute for an ISO 17025 certified report. Dively Scale Co., Inc. certifies that the above listed equipment pass or fail based on a comparison between displayed results and the applicable tolerance. Our estimate of Measurement Uncertainty is not considered in these pass / fail conclusions and can be provided upon request for an additional fee. Our test methods comply with National and International Standards (e.g. N.I.S.T. Handbook 44 where applicable) or by validated customer specified methods. The test weights used are certified to comply with standard masses, traceable to N.I.S.T. via a ISO 17025 accredited or N.I.S.T. approved labs. Any discrepancies beyond acceptable tolerances, in the aforementioned equipment, are reported to the appropriate facility's manager for authorization of further repairs and/or replacement as necessary. This report shall not be reproduced except in full without expressed written consent from Dively Scale Co., Inc.



DIVELY SCALE CO., INC.

4012 Alameda Ave., El Paso, TX 79905-3799
 (915) 533-6712 or (888) 383-6744 FAX 533-6755
 1607 4th St. N.W., Albuquerque, NM, 87102-1622
 (505) 243-2189 FAX 242-2485

Calibration Traceability Certificate

Work Order No.:

9845

Page 4 of 4

Customer:	Double Tree Hotel	Next Calibration Due	
Address:	701 Marquette Ave NW	Work Instruction:	TF N/A
City, State & Zip:	Albuquerque, NM 87102	Environmental Conditions	
Service Location (If Different):		Temperature (°F):	N/A
Calibration Date:	9-25-14	Relative Humidity (%):	N/A
Calibration Interval:	As needed	Any Adverse Conditions (Yes/No)?	NO
Indicator Mfg:	Total Comp	Model #:	T300E
Base Mfg:		Serial #:	507001100027
Capacity:	400 lb	Readability:	0.02
		Divisions:	20000
TOLERANCE CODE		Handbook 44:	11.1
		Customer Specified:	
		Other (See Comments):	

REPEATABILITY

(Pass/Fail)
 As Found: Pass
 As Left: Pass

Section, Corner, or Side-To-Side Test

INDICATOR	Position	As Found		Position	As Left	
		Reading	Error		Reading	Error
Platform Layout 3 2 4 1	1	0.00		1		
	2	0.00	Error 0.00	2	BP	Error
	3	0.00	Tolerance +/- 0.02	3		Tolerance +/-
	4	0.00	Pass / Fail Pass	4		Pass / Fail
	5		Load Used: 100.00	5		Load Used:
	6			6		

Load Test

Test Weight Applied	As Found Reading	Error	Tolerance +/-	Pass / Fail	As Left Reading	Pass / Fail
ZERO	0.00	0.00	0.00	Pass		
100.00	100.00	0.00	0.02	Pass		
200.00	200.00	0.00	0.02	Pass	BP	
300.00	300.00	0.00	0.04	Pass		
400.00	400.00	0.00	0.04	Pass		
200.00	200.00	0.00	0.02	Pass		
0.00	0.00	0.00	0.02	Pass		

S/N of N.I.S.T. Traceable Test Weights Used: A60 - A65

N.I.S.T. Reference Number

13-187

Comments:

Technician

BAJ. DM

State Reg. No.

195

Date

9-25-14

FOR QUESTIONS AND CLARIFICATIONS PLEASE CONTACT THE SERVICE MANAGER

This certificate is NOT ISO 17025 compliant and should not be used as a substitute for an ISO 17025 certified report. Dively Scale Co., Inc. certifies that the above listed equipment pass or fail based on a comparison between displayed results and the applicable tolerance. Our estimate of Measurement Uncertainty is not considered in these pass / fail conclusions and can be provided upon request for an additional fee. Our test methods comply with National and International Standards (e.g. N.I.S.T. Handbook 44 where applicable) or by validated customer specified methods. The test weights used are certified to comply with standard masses, traceable to N.I.S.T. via a ISO 17025 accredited or N.I.S.T. approved labs. Any discrepancies beyond acceptable tolerances, in the aforementioned equipment, are reported to the appropriate facility's manager for authorization of further repairs and/or replacement as necessary. This report shall not be reproduced except in full without expressed written consent from Dively Scale Co., Inc.